

checkCIF/PLATON report

Structure factors have been supplied for datablock(s) pov2

THIS REPORT IS FOR GUIDANCE ONLY. IF USED AS PART OF A REVIEW PROCEDURE FOR PUBLICATION, IT SHOULD NOT REPLACE THE EXPERTISE OF AN EXPERIENCED CRYSTALLOGRAPHIC REFEREE.

No syntax errors found. CIF dictionary Interpreting this report

Datablock: pov2

Bond precision:	O- B = 0.0025 A	Wavelength=0.71073
Cell:	a=16.2308 (3)	b=16.2308 (3) c=7.4676 (1)
	alpha=90	beta=90 gamma=120
Temperature:	293 K	
	Calculated	Reported
Volume	1703.69 (7)	1703.69 (7)
Space group	R 3 m	R 3 m :H
Hall group	R 3 -2"	R 3 -2"
Moiety formula	Al _{15.63} B ₉ Fe _{20.36} H ₉ Mg O ₉₃ Si ₁₈ , 0.699 (K), 2.301 (Na)	?
Sum formula	Al _{15.63} B ₉ Fe _{20.36} H ₉ K _{0.70} Mg Na _{2.30} O ₉₃ Si ₁₈	H ₃ Al _{11.88} B ₃ Fe _{6.78} K _{0.23} Mg _{0.34} Na _{0.77} O ₃₁ Si ₆
Mr	3493.72	1164.47
Dx, g cm ⁻³	3.405	3.405
Z	1	3
Mu (mm ⁻¹)	4.823	4.823
F000	1703.2	1889.0
F000'	1713.13	
h, k, lmax	27, 27, 12	27, 27, 12
Nref	2149 [1081]	2067
Tmin, Tmax	0.311, 0.381	0.585, 0.747
Tmin'	0.288	

Correction method= # Reported T Limits: Tmin=0.585 Tmax=0.747

AbsCorr = MULTII-SCAN

Data completeness= 1.91/0.96

Theta (max)= 37.379

R(reflections)= 0.0147(2028)

wR2(reflections)=
0.0283(2067)

S = 1.062

Npar= 94

The following ALERTS were generated. Each ALERT has the format

test-name_ALERT_alert-type_alert-level.

Click on the hyperlinks for more details of the test.

Alert level B

PLAT927_ALERT_1_B Reported and Calculated wR2 Differ by -0.0066 Check
PLAT928_ALERT_1_B Reported and Calculated S value Differ by . -0.247 Check

Alert level C

PLAT041_ALERT_1_C Calc. and Reported SumFormula Strings Differ Please Check
PLAT068_ALERT_1_C Reported F000 Differs from Calcd (or Missing)... Please Check
PLAT077_ALERT_4_C Unitcell Contains Non-integer Number of Atoms .. Please Check
PLAT926_ALERT_1_C Reported and Calculated R1 Differ by -0.0017 Check
PLAT975_ALERT_2_C Check Calcd Resid. Dens. 0.57Ang From O3 . 0.40 eA-3

Alert level G

CELLZ01_ALERT_1_G Difference between formula and atom_site contents detected.
CELLZ01_ALERT_1_G ALERT: check formula stoichiometry or atom site occupancies.

From the CIF: _cell_formula_units_Z 3

From the CIF: _chemical_formula_sum H3 Al1.88 B3 Fe6.78 K0.23 Mg0.34 N

TEST: Compare cell contents of formula and atom_site data

atom	Z*formula	cif sites	diff
H	9.00	9.00	0.00
Al	5.64	5.63	0.01
B	9.00	9.00	0.00
Fe	20.34	20.36	-0.02
K	0.69	0.70	-0.01
Mg	1.02	1.01	0.01
Na	2.31	2.30	0.01
O	93.00	93.00	-0.00
Si	18.00	18.00	0.00

PLAT002_ALERT_2_G Number of Distance or Angle Restraints on AtSite 2 Note
PLAT004_ALERT_5_G Polymeric Structure Found with Maximum Dimension 1 Info
PLAT017_ALERT_1_G Check Scattering Type Consistency of T as SI
PLAT045_ALERT_1_G Calculated and Reported Z Differ by a Factor ... 0.333 Check
PLAT168_ALERT_4_G The CIF-Embedded .res File Contains EXYZ Records 3 Report
PLAT171_ALERT_4_G The CIF-Embedded .res File Contains EADP Records 3 Report
PLAT172_ALERT_4_G The CIF-Embedded .res File Contains DFIX Records 1 Report
PLAT199_ALERT_1_G Reported _cell_measurement_temperature (K) 293 Check
PLAT200_ALERT_1_G Reported _diffrn_ambient_temperature (K) 293 Check
PLAT301_ALERT_3_G Main Residue Disorder(Resd 1) 19% Note
PLAT302_ALERT_4_G Anion/Solvent/Minor-Residue Disorder (Resd 2) 100% Note
PLAT302_ALERT_4_G Anion/Solvent/Minor-Residue Disorder (Resd 3) 100% Note
PLAT396_ALERT_2_G Deviating Si-O-Si Angle From 150 for O5 . 133.4 Degree
PLAT720_ALERT_4_G Number of Unusual/Non-Standard Labels 5 Note
PLAT883_ALERT_1_G No Info/Value for _atom_sites_solution_primary . Please Do !
PLAT912_ALERT_4_G Missing # of FCF Reflections Above STh/L= 0.600 23 Note

PLAT982_ALERT_1_G The O-f' = 0.0080 Deviates from IT-value = 0.0106 Check

0 **ALERT level A** = Most likely a serious problem - resolve or explain
2 **ALERT level B** = A potentially serious problem, consider carefully
5 **ALERT level C** = Check. Ensure it is not caused by an omission or oversight
19 **ALERT level G** = General information/check it is not something unexpected

13 ALERT type 1 CIF construction/syntax error, inconsistent or missing data
3 ALERT type 2 Indicator that the structure model may be wrong or deficient
1 ALERT type 3 Indicator that the structure quality may be low
8 ALERT type 4 Improvement, methodology, query or suggestion
1 ALERT type 5 Informative message, check

Validation response form

Please find below a validation response form (VRF) that can be filled in and pasted into your CIF.

```
# start Validation Reply Form
_vrf_PLAT927_pov2
;
PROBLEM: Reported and Calculated wR2 Differ by ..... -0.0066 Check
RESPONSE: ...
;
_vrf_PLAT928_pov2
;
PROBLEM: Reported and Calculated S value Differ by . -0.247 Check
RESPONSE: ...
;
_vrf_PLAT041_pov2
;
PROBLEM: Calc. and Reported SumFormula Strings Differ Please Check
RESPONSE: ...
;
_vrf_PLAT068_pov2
;
PROBLEM: Reported F000 Differs from Calcd (or Missing)... Please Check
RESPONSE: ...
;
_vrf_PLAT077_pov2
;
PROBLEM: Unitcell Contains Non-integer Number of Atoms .. Please Check
RESPONSE: ...
;
_vrf_PLAT926_pov2
;
PROBLEM: Reported and Calculated R1 Differ by ..... -0.0017 Check
RESPONSE: ...
;
_vrf_PLAT975_pov2
;
PROBLEM: Check Calcd Resid. Dens. 0.57Ang From O3 . 0.40 eA-3
RESPONSE: ...
;
# end Validation Reply Form
```

It is advisable to attempt to resolve as many as possible of the alerts in all categories. Often the minor alerts point to easily fixed oversights, errors and omissions in your CIF or refinement strategy, so attention to these fine details can be worthwhile. In order to resolve some of the more serious problems it may be necessary to carry out additional measurements or structure refinements. However, the purpose of your study may justify the reported deviations and the more serious of these should normally be commented upon in the discussion or experimental section of a paper or in the "special_details" fields of the CIF. checkCIF was carefully designed to identify outliers and unusual parameters, but every test has its limitations and alerts that are not important in a particular case may appear. Conversely, the absence of alerts does not guarantee there are no aspects of the results needing attention. It is up to the individual to critically assess their own results and, if necessary, seek expert advice.

Publication of your CIF in IUCr journals

A basic structural check has been run on your CIF. These basic checks will be run on all CIFs submitted for publication in IUCr journals (*Acta Crystallographica*, *Journal of Applied Crystallography*, *Journal of Synchrotron Radiation*); however, if you intend to submit to *Acta Crystallographica Section C* or *E* or *IUCrData*, you should make sure that full publication checks are run on the final version of your CIF prior to submission.

Publication of your CIF in other journals

Please refer to the *Notes for Authors* of the relevant journal for any special instructions relating to CIF submission.

